

Certificate of Conformance

Date: 10/10/2025
Part Name: Female ENFit® to Female Luer Lock Adapter -
RD Part Number: EL02
Description of Part: Connects male ENFit® (e.g. Rx Crush Pouch) to male Luer syringe.
RD SKUs: EL02-5, EL02-100, EL02-1000, EL02-4000
UoMs: 5/Sample, 100/Pack, 1000/Box
UoM Packaging Contents: 5 Pieces, 100 Pieces, 10 Packs of 100 Pieces
UoM Dims LxWxH (cm): EL02-5: 15.2 x 10.2 x 0.6, EL02-100: 17.8 x 11.4 x 2.5, EL02-1000: 20.3 x 15.2 x 15.2

UoM Weight (g): EL02-5: 4, EL02-100: 76, EL02-1000: 762
GTINs: 10850006611123, 00850006611126, 20850006611120
Lot: 310132
Expiration Date: 3/17/2030
Material: ABS, Colorant

Made In:
HTS: 9018.31.0090
NMFC: 47550
EMDN: A0799
GMDN: 64470

FDA Establishment #, FDA Listing #, Product Code, Classification: 3013083631, D276971, PIO, Class 2

EU/CE Actor ID, Basic UDI-DI, Class: US-MF-000017382, 0850006611AD67, Class I

EU REACH 1907/2006: Based on our knowledge, including information from third parties, article does not contain Substances of Very High Concern (SVHC) above 0.1% by weight.

EU MDR 2017/745, Section 10.4: Based on our knowledge, including information from third parties, article does not contain prohibited materials.

RoHS 2015/863/EU Annex II: Based on our knowledge, including information from third parties, article does not contain Substances of Very High Concern (SVHC) above 0.1% by weight.

California Prop 65: Article does not contain materials on the California Proposition 65 list of prohibited substances.
Standards Compliance: ISO 80369-3

We hereby certify that the materials used in the production of the above items will not be changed without prior notification to our customers.

We hereby certify that the aforementioned part has been inspected and released by our Quality department.



Aaron Collett
Quality Manager, Resolve Designs
Aaron@RxCrush.com
+01.877.343.2783 x 803